

MAINTENANCE BOOK

Ortho Dept

17-8-2020

Repairing Auto clare & replace the
Heater coil.

Dept of Prost

Repairing two electrical circuits
supply connection. #

HEAD OF THE DEPARTMENT
Dept. of Prosthodontics
S.J.M. Dental College & Hospital
CHITRADURGA-577501.

Dept of Pedo

Repair chair light Helogen Ballp.
working conditions.

Leotale
PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA.

PROFESSOR & H.O.D
Dept. of Pedodontics & Preventive Dentistry
S.J.M. DENTAL COLLEGE & HOSPITAL
CHITRADURGA

Slip No - Dept


Date - 9-8-2020

Repair electrical chair up & down
connection.



Period Dept


Repair scales, Air connection & replace
Air tube & separator for the
water connection.




Head of Periodontics
College & Hospital
CHITRADURGA.

Conservative Dept

Repair electrical chair sections
connection working condition.


PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA.


HEAD OF THE DEPARTMENT OF
CONSERVATIVE DENTISTRY & PERIODONTICS
S.J.M. Dental College & Hospital
CHITRADURGA.

Dept of Prost

12/8/2020

Repair Chair water connection spitter
of tubed. working in good
condition.

HEAD OF THE DEPARTMENT
Dept of Prost
S.J.M. Dental College & Hospital
CHITRADURGA-577501

Dept of Pedo

Repair Chair light Helogen Bulb
connection.

M Peddy

PROFESSOR & H.O.D
Dept of Pedodontics & Preventive Dentistry
S.J.M. Dental College & Hospital
CHITRADURGA

Dental Material Lab

Repair Heating water Bulb
Replace the Sapper wire.

Chopra
PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA.

HEAD OF THE DEPARTMENT
Dept of Prost
S.J.M. Dental College & Hospital
CHITRADURGA-577501


Dept P.cio Lab

16-8-2020

Repair Pantan Head. scupper connection.
at micro metal connection.

Dept of oral Surgery

Repair Chair Light Helogen Bulb
connection. working condition.



PROFESSOR & HEAD
Dept of Oral and Maxillofacial Surgery
S.J.S. Dental College & Hospital
Chitradurga

Dept of Oral Patho.

Repair microscope unit.
working condition

Dept. of PCP

Georgina
PRINCIPAL
S.J.S. D.C. & H. CHITRADURGA.

HEAD OF THE DEPARTMENT
Dept of Prosthodontics
S.J.S. Dental College & Hospital
CHITRADURGA-577501.



Conservative Dept

20-8-2020

Repair pulp testing and Re place
the Bathys working condition

HEAD OF THE DEPARTMENT OF
CONSERVATIVE DENTISTRY & ENDODONTICS
S.J.M. Dental College & Hospital
CHITRADURGA.

Dept of Perio

Repair Sealer unit water connection
& fix connection working condition

Head of Periodontics
S.J.M. Dental College & Hospital
CHITRADURGA.

Dept. Ortho


Repair Sealer unit
in ortho Dept.

PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA.

Conservative Dept

23-8-2020

Repair Air-Lag Hand piece
Two no.


HEAD OF THE DEPARTMENT OF
CONSERVATIVE DENTISTRY & ENDODONTICS
S.J.M. Dental College & Hospital
CHITRADURGA.


Prosth. Dept.


Repair Chair Supply connection
& Replace the 3 core wire
Six meter

HEAD OF THE DEPARTMENT
S.J.M. Dental College & Hospital
CHITRADURGA-577501.

Pedo Dept.

Repair Air-Lag Hand piece
In working condition.





PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA.

PROFESSOR & H.O.D.
Dept. of Preventive Dentistry
& Hospital

Dept of Oral Patho

26-8-2020

Repair microscopes. Unit three are.
light- and supply connections

Dept of Prosth

Repair Air tag Hand piece
working in good condition.

HEAD OF THE DEPARTMENT
Dept. of Prosthodontics
S.J.M. Dental College & Hospital
CHITRADURGA-577501.

Dept. of Perio.

Repair chair Back Rest and up
down condition. Working in
good condition.

Dr. H. D. Sumas
Dept. of Periodontics
College & Hospital
CHITRADURGA.

CHSE

U. M. P. S.
PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA.

Dept of Pedo

28-8-2020

Repair Chair Back Rest &
Helogen light connections.

M Peddy

PROFESSOR
Dept. of Pedodontics & Preventive Dentistry
S.J.M.C. Dental College & Hospital
CHITRADURGA

Dept of oral medicine

Repair Chair light connections
Re place Helogen Bulb.

L S

Head of the Dept
Dept. of Oral Medicine
S.J.M.C. & H
CHITRADURGA

Prosth Lab

Repair panting Head unit and
supply connection.

[Signature]

HEAD OF THE DEPARTMENT
Dept. of Prosthodontics
S.J.M.C. Dental College & Hospital
CHITRADURGA-577501.


[Signature]

PRINCIPAL
S.J.M.C. & H. CHITRADURGA

conservative Dept.

1-9-2020


Repair Chair Air/Water connections
Air connections for the Suction
working condition


HEAD OF THE DEPARTMENT OF
CONSERVATIVE DENTISTRY & ENDO DONTICS
S.J.M. Dental College & Hospital
CHITRADURGA.

Dept Prosth

Repair Chair water supply
Spitton & tumbler connection.

HEAD OF THE DEPARTMENT
Dept. of Prosthodontics
S.J.M. Dental College & Hospital
CHITRADURGA-577501.



PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA.

Dept of Perio

12-9-2020

Re place the Helogen Bulb.
Light connection, working condition.

Dr HOD Jumb

Periodont
& Hospital
PCA.

Dept of ortho.

Repair 3 way syringe water &
Air connection.

Dept of Prost Lab

Repair micro metal cond. working
good condition.

HEAD OF THE DEPARTMENT
Dept of Prosthodontics
S.J.M. Dental College & Hospital
CHITRADURGA-577501.


Leone me

PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA.

Dept of P.C.O. Lab

15-9-2023

Repair Pentan Head 12 no. And
Jaw Setting. working good
condition.


HEAD OF THE DEPARTMENT OF
CONSERVATIVE DENTISTRY & ENDODONTICS
S.J.M. Dental College & Hospital
CHITRADURGA.

Dept of ortho.

Repair 3 way spraying water
Air connection, working in
good condition.

Air Compressor Room

Repair Air compressor metal
supply connection.

Leompie
PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA.

ELECTRICAL MAINTN

20/8/20 Population + social.
in the house report out C.S.

31/5/20 Am Fe close 2 K Lines
Co Ce Final
Shipping July

4/9/20 Also one Fine table
Final

leopard me

PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA.

17/9/20 A. E. H. ...
... ..

[Handwritten signature]

23/9/20 Store Room :-

Install the Three No Tub Light
and Repair the fan and
two No Sp Switch Change

29/9/20 Ceramic Lab :-

Change the Dabhal pot Switch
and Carry over Switch

[Handwritten signature]

6/10/20 Ortho Lab :-

Change the fan Regulator
and Service the fan

PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA
PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA.

[Handwritten signature]

13/10/20

PCC Lab :-

Wiring the ~~chair~~ chair
and check the supply and
instal the tube light

18/10/20

Cond :-

Change the switch to
box and D.P Switch Two No
fan check the Earthing Line

22/10/20

Library :-

Instal the wall
mounting fan, one ceiling fan
Led tube light.

HEAT
CON
S.I.A

Ching

4/11/20

Pop office :-

Computer wiring, change the
tube light, one no wall
mounting fan

9/11/20

Principal office

Change the pop light.

Repair the A.C

Check the ceiling bed

Leomp me

PRINCIPAL

S.J.M.D.C.&H. CHITRADURGA.

AD-CASTRO H.C.C.M.L.S

15/11/20 examination Room :-
Service the stand fan
change the tube light
Jef

19/11/20 power room :-
change the 415 V. fuse 3 no
Jef

25/11/20 period :-
one no ceiling fan instal
Two No Led tube light. instal
check the chair wiring

29/11/20 oral pathology :-
New earthing wiring. Chinn

HEAD OF THE DEPARTMENT
Department Of Oral Pathology
S. J. M. Dental College & Hospital
CHITRADURGA- 577501

20/12/20 Dental material Room :-
check the grinding machine
check the gas pipe line

Ortho :-

PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA

0009pm
PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA.

1/12/20

Ortho ~~DPT~~ DPT :-

Switch box change
Change the sterilizer coil
Repair the Auto Clave

D. W. Annamalai

DEPT. OF ORTHODONTICS
S. J. M. Dental College & Hospital
CHITRADURGA.

7/12/20

Pedo DPT :-

HOD Room wiring change
fan Re winding
tube light instal

M. Peddath
PROFESSOR & HEAD
Dept. of Pedodontics & Preventive Dentistry
S.J.M. Dental Coll. & Hospital
CHITRADURGA

18/12/10

OMFS DPT :-

Repair the Chair Light
Chair cable change

PROFESSOR & HEAD
PROFESSOR & HEAD
Dept. of Oral And Maxillofacial Surgery
S.J.M. Dental College & Hospital
Chitradurga

20/12/10

OMRD DPT :-


Auto Clave coil change
Wire Card repair

LEONARDO
PRINCIPAL
S.J.M.D.C.&H. CHITRADURGA.

22/12/20

counter.


fan wiring Led tube light
2 no pop light. 2 no
check the all ceiling fan.



24/12/20

College. ~~labor~~ park

check the all tube light
and Street.




31/12/20

College clamps.

Cable change.


instal the D.P switch.



4/1/21

ground floor corridor.

check the wiring, and instal
the tube light 6 no



12/1/21

2nd floor corridor.

PRINCIPAL
S.J.M.D.C.&H.CHITRADURGA


all light paint checking.
complete

PRINCIPAL
S.J.M.D.C.&H.CHITRADURGA

18/1/21

2nd floor corridor:-

Change the wiring and and
Install the light point.




28/1/21

Aditorium room:-

3 No pop light


Repair the wall mounting fan



OP Lab. :-

10/2/21

Change the top switch
and D.P Switch.




19/2/21

AV hall :-

Check the lighting point

2 no ceiling fan.

Wiring change.



PRINCIPAL

S.J.M.D.C.&H. CHITRADURGA.

1/3/21

QAAC ROOM

Repair the stand fan
change the CFL bulb.

18/3/21

Watch man room
change the tubular light.

PROFESSOR & HEAD
Dept. of Oral And Maxillofacial Surgery
S.J.M. Dental College & Hospital
Chitradurga

[Signature]

26/3/21

VIP room

fan switch change.
instal the 2 no led tubelight.

[Signature]

9/4/21

Bodded room

change the pop light
Service the AC point.

HEAD OF THE DEPARTMENT

12/11/21

CHITRADURGA

CHITRADURGA

[Signature]

PRINCIPAL

S.J.M.D.C.&H. CHITRADURGA.

PRINCIPAL

S.J.M.D.C.&H. CHITRADURGA

12/17/24

PHD Dept

4 No Spswitch change

6 No fan Regulator

2 No Led tube

Pvc wire 10 meter wire

check the chair light.

16/6/21

PS TO principal Room

Computer wiring

2 No Led tube

Change the switches

Sneha K.

14/7/24

prosthodontics Dept

Auto clare coil change

change fan change 2 No

pop light. instal

led. tube light change

Chair wiring.

HEAD OF THE DEPARTMENT

Dept. of Prosthodontics

S.J.M. Dental College & Hospital

CHITRADURGA-577501.

Leokhep m me

PRINCIPAL

S.J.M.D.C.&H. CHITRADURGA.